

3M™ Petrifilm™ *E. coli*/Coliform Count Plates

Product Information

MEDIUM:	3M Petrifilm <i>E. coli</i> /Coliform Count Plates (6404 /6414/6444)								
ISO/BSI CERTIFICATE OF REGISTRATION NUMBER:	Manufactured at Brookings, SD, USA (ISO 9001:2008, FM 14552)								
DATE OF EXPIRATION/ LOT NUMBER:	Expiry and lot number indicated on each package. Lot number indicated on each plate.								
FORMULATION:	Modified Violet Red Bile (VRB) nutrients, cold-water soluble gel, an indicator of glucuronidase activity, 5-bromo-4-chloro-3-indolyl-D-glucuronide, and tetrazolium indicator								
METHOD OF PREPARATION:	Nutrients and gels coated onto film. For use, hydrate with one mL aqueous sample or dilution of sample. See product instructions for detailed instructions.								
CONTAMINATION CHECK:	96 randomized plates per batch tested Incubated at 32°C for 24 and 48 hours Utilizes statistical sampling plan								
EFFICACY CHECK:	Complement of organisms tested includes, among others: <table><tr><td><u>Organism</u></td><td><u>Acceptable Batch</u></td></tr><tr><td><i>Escherichia coli</i> ATCC 51813</td><td>Quality Test Statistic $\geq -3^*$</td></tr><tr><td><i>Enterobacter amnigenus</i> ATCC 51816</td><td>Quality Test Statistic $\geq -3^*$</td></tr><tr><td><i>Enterococcus faecalis</i> ATCC 14506</td><td>No Growth</td></tr></table> <p>*Result must be greater than or equal to 3 standard deviations below the mean count on reference agar</p>	<u>Organism</u>	<u>Acceptable Batch</u>	<i>Escherichia coli</i> ATCC 51813	Quality Test Statistic $\geq -3^*$	<i>Enterobacter amnigenus</i> ATCC 51816	Quality Test Statistic $\geq -3^*$	<i>Enterococcus faecalis</i> ATCC 14506	No Growth
<u>Organism</u>	<u>Acceptable Batch</u>								
<i>Escherichia coli</i> ATCC 51813	Quality Test Statistic $\geq -3^*$								
<i>Enterobacter amnigenus</i> ATCC 51816	Quality Test Statistic $\geq -3^*$								
<i>Enterococcus faecalis</i> ATCC 14506	No Growth								
PACKAGING:	Pack size: 25 plates per foil pouch Film grade: Plastic foil laminate Seal integrity check: Pressure Decay test performed								
MEASUREMENT SYSTEMS CALIBRATION AND TRACEABILITY:	Incubator temperature 3M internal calibration Minimum calibration once per year for all equipment								
MEDIA QUALITY STATEMENT:	Certificate of Analysis available on 3m.com/foodsafety								
SHELF LIFE:	18 months from date of manufacture								
STORAGE CONDITIONS:	Store at temperatures less than or equal to 8°C								

**HACCP
PLUS+**

02 9099 5988
info@haccplus.com.au
www.haccplus.com.au